

Calibrating Unit New Standard Solution Report

Equipment

Alcotest 7110 MKIII-C		Serial No.: ARTL-0009
Location: DUNELLEN POLICE DEPT.		
Calibration File No.: 01264	Calib. Date: 11/30/2018	Calib. No.: 00039
Certification File No.: 01265	Cert. Date: 11/30/2018	Cert. No.: 00030
Linearity File No.: 01266	Lin. Date: 11/30/2018	Lin. No.: 00031
Solution File No.: 01275	Soln. Date: 01/27/2019	Soln. No.: 00251
Sequential File No.: 01275	File Date: 01/27/2019	

Calibrating Unit: WET	Model No.: CU-34	Serial No.: DDUM S3-0218
Control Solution %: 0.100%		Expires: 09/17/2020
Solution Control Lot: 18290		Bottle No.: 1375

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	20:50S		
Control 1 EC	0.100%	20:51S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	20:51S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:52S		
Control 2 EC	0.100%	20:52S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	20:52S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:53S		
Control 3 EC	0.101%	20:54S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	20:54S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:54S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.


Changed By:

Last Name: DOHERTY

First Name: JOHN

MI: J

Signature: _____



* 30

Badge No.: 30

Date: 01/27/2019