

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARTL-0009
Location:	DUNELLEN POLICE DEPT.	Calib. No.:	00041
Calibration File No.:	01331	Cert. Date:	09/20/2019
Certification File No.:	01332	Lin. Date:	09/20/2019
Linearity File No.:	01333	Soln. Date:	02/14/2020
Solution File No.:	01373	File Date:	02/14/2020
Sequential File No.:	01373		
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUM S3-0213
Solution Control Lot:	19240	Expires:	09/09/2021
		Bottle No.:	0995

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	04:09S		
Control 1 EC	0.098%	04:09S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	04:09S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:10S		
Control 2 EC	0.098%	04:11S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	04:11S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:11S		
Control 3 EC	0.098%	04:12S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.098%	04:12S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:13S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: ROBBINS

First Name: BRIAN

MI: E

Signature: _____

Sgt. B. Robbins #41

Badge No.: 41

Date: 02/14/2020