

Calibrating Unit New Standard Solution Report

Equipment

Alcotest 7110 MKIII-C
 Location: DUNELLEN POLICE DEPT. Serial No.: ARTL-0009
 Calibration File No.: 01529 Calib. Date: 03/15/2022 Calib. No.: 00047
 Certification File No.: 01530 Cert. Date: 03/15/2022 Cert. No.: 00038
 Linearity File No.: 01531 Lin. Date: 03/15/2022 Lin. No.: 00039
 Solution File No.: 01541 Soln. Date: 04/24/2022 Soln. No.: 00294
 Sequential File No.: 01541 File Date: 04/24/2022

Calibrating Unit: WET Model No.: CU-34 Serial No.: DDUM S3-0218
 Control Solution %: 0.100% Expires: 03/17/2023
 Solution Control Lot: 21090 Bottle No.: 0952

Function	Result	Time	Temperature	Comment(s) or Error(s)
	%BAC	HH:MM	Simulator (°C)	
Ambient Air Blank	0.000%	21:45D		
Control 1 EC	0.099%	21:46D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	21:46D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:46D		
Control 2 EC	0.099%	21:47D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	21:47D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:48D		
Control 3 EC	0.098%	21:48D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.100%	21:48D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:49D		

All tests within acceptable tolerance.


On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: DELBUONO

First Name: JEFFREY

MI: M

Signature: 

Badge No.: 32

Date: 04/24/2022


DEPARTMENT OF
Traffic and Public Safety
This is for receiving that


JEFFREY DELBUONO
DUNELLEN BOROUGH POLICE

ALCOTEST 7110 MKIII-C







IN QUALIFIED AND COMPETENT TO CONDUCT OPERATIONS AND ANALYSES PURSUANT TO CHAPTER 14C OF THE LAWS OF 1966 IN THE OPERATION OF THE ALCOTEST 7110 MKIII-C A METHOD TO DETERMINE INTOXICATION GIVEN UNDER MY HAND AT TRENTON, NEW JERSEY

Mid DAY OF July
TWO THOUSAND AND NINE

 SUPERINTENDENT
NEW JERSEY STATE POLICE

 ATTORNEY GENERAL
STATE OF NEW JERSEY

ORIGINAL COURSE DATES

DATE	Refresher Course PLACE	INSTRUCTOR
2-24-11	SAVERVILLE PD	
5-7-13	SAVERVILLE	
1-3-15	SAVERVILLE PD	
7-8-17	SAVERVILLE PD	
7-16-19	SAVERVILLE PD	
10-14-21	Saverville PD	

Certificate of Training
This is to certify that

Jeffrey Delbuono

has successfully completed practitioner training in the use of
Horizontal Gaze Nystagmus
Improved Sobriety Test Procedures
as set forth by the National Highway Traffic Safety Administration

