

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.:	ARTL-0009
Location:	DUNELLEN POLICE DEPT.		
Calibration File No.:	01529	Calib. Date:	03/15/2022
Certification File No.:	01530	Cert. Date:	03/15/2022
Linearity File No.:	01531	Lin. Date:	03/15/2022
Solution File No.:	01550	Soln. Date:	05/21/2022
Sequential File No.:	01550	File Date:	05/21/2022
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%		Serial No.:
Solution Control Lot:	21090		DDUM S3-0218
			Expires: 03/17/2023
			Bottle No.: 1176

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	02:48D		
Control 1 EC	0.100%	02:48D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	02:48D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:49D		
Control 2 EC	0.100%	02:50D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	02:50D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:50D		
Control 3 EC	0.099%	02:51D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	02:51D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	02:52D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

Last Name: DELBUONO

First Name: JEFFREY

MI: M

Signature: \_\_\_\_\_



Badge No.: 32

Date: 05/21/2022

**DEPARTMENT OF  
Traffic and Public Safety**  
*This is to certify that*

**JEFFREY DELBUONO**  
**DUNELLEN BOROUGH POLICE**

IS QUALIFIED AND COMPETENT TO CONDUCT OPERATIONAL BREATH ANALYSES PURSUANT TO CHAPTER 142 OF THE LAWS OF 1966 IN THE OPERATION OF THE **ALCOTEST 7110 MKIII-C**

GIVEN UNDER MY HAND AT TRENTON, NEW JERSEY, THE \_\_\_\_\_ and \_\_\_\_\_ DAY OF July  
TWO THOUSAND AND NINE

**Certificate of Training**  
*This is to certify that*

**Jeffrey Delbuono**

has successfully completed practitioner training in the use of **Horizontal Gaze Nystagmus** and **Improved Sobriety Test Procedures** as set forth by the National Highway Traffic Safety Administration

\_\_\_\_\_  
SUPERVISOR

\_\_\_\_\_  
INSTRUCTOR

DATE	PLACE	INSTRUCTOR
1. 2-24-11	SPARVILLE PD	
2. 5-7-13	SAYREVILLE	
3. 1-13-15	SAYREVILLE PD	
4. 7/6/17	SAYREVILLE PD	
5. 11/26/17	Sayreville PD	
6. 10/14/21	Sayreville PD	
7.		