

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARTL-0009
Location:	DUNELLEN POLICE DEPT.	
Calibration File No.:	01529	Calib. Date: 03/15/2022
Certification File No.:	01530	Calib. No.: 00047
Linearity File No.:	01531	Cert. Date: 03/15/2022
Solution File No.:	01559	Cert. No.: 00038
Sequential File No.:	01559	Lin. Date: 03/15/2022
		Lin. No.: 00039
		Soln. Date: 06/19/2022
		Soln. No.: 00296
		File Date: 06/19/2022
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUM S3-0218
Solution Control Lot:	21090	Expires: 03/17/2023
		Bottle No.: 0131

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	21:14D		
Control 1 EC	0.100%	21:15D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.099%	21:15D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:16D		
Control 2 EC	0.099%	21:16D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	21:16D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:17D		
Control 3 EC	0.099%	21:18D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	21:18D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:18D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: DELBUONO

First Name: JEFFREY

MI: M

Signature: *Jeffrey Delbuono*

Badge No.: 32

Date: 06/19/2022

Certificate of Training

I wish to certify that

Jeffrey Delbuono

has successfully completed practitioner training in the use of

Horizontal Gaze Nystagmus

Improved Sobriety Test Procedures

as set forth by the National Highway Traffic Safety Administration

Instructor: *Jeffrey Delbuono*

Division of Highway Traffic Safety

ORIGINAL COURSE DATES

DATE	Refresher Course PLACE	INSTRUCTOR
1. 2-24-11	SAYREVILLE PD	<i>[Signature]</i>
2. 5-7-13	JAYREVILLE	<i>[Signature]</i>
3. 1-3-15	SAYREVILLE PD	<i>[Signature]</i>
4. 7-15-17	SAYREVILLE PD	<i>[Signature]</i>
5. 7-15-19	SAYREVILLE PD	<i>[Signature]</i>
6. 10-14-21	SAYREVILLE PD	<i>[Signature]</i>
7.		
8.		

DEPARTMENT OF
Transportation and Public Safety

This is to certify that

JEFFREY DELBUONO

DUNELLEN BOROUGH POLICE

IS QUALIFIED AND COMBENT TO CONDUCT THE ABOVE TEST PURSUANT TO CHAPTER 148 OF THE LAWS OF 1966 IN THE OPERATION OF THE ALCO TEST 7110 MKIII-C

A METHOD TO DETERMINE INTOXICATION

GIVEN UNDER MY HAND AT TRENTON, NEW JERSEY, this 19 DAY OF July

TWO THOUSAND AND NINE

[Signature]
SUPERINTENDENT
NEW JERSEY STATE POLICE

[Signature]
ATTORNEY GENERAL
STATE OF NEW JERSEY