

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARTL-0009
Location:	DUNELLEN POLICE DEPT.	
Calibration File No.:	01482	Calib. Date: 05/12/2021
Certification File No.:	01483	Calib. No.: 00046
Linearity File No.:	01484	Cert. Date: 05/12/2021
Solution File No.:	01498	Cert. No.: 00037
Sequential File No.:	01498	Lin. Date: 05/12/2021
		Lin. No.: 00038
		Soln. Date: 06/11/2021
		Soln. No.: 00287
		File Date: 06/11/2021
Calibrating Unit:	WET	Model No.: CU-34
Control Solution %:	0.100%	Serial No.: DDUM S3-0218
Solution Control Lot:	21090	Expires: 03/17/2023
		Bottle No.: 1173

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	06:09D		
Control 1 EC	0.102%	06:10D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	06:10D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:11D		
Control 2 EC	0.101%	06:11D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	06:11D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:12D		
Control 3 EC	0.101%	06:13D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	06:13D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	06:14D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: NIERODZIK

First Name: MARCIN

MI: -

Signature: *Marcin Nierodzik #46*

Badge No.: 46

Date: 06/11/2021

Certificate of Training
This is to certify that

Marcin Nierodzik

has successfully completed practitioner training in the use of
Horizontal Gaze Nystagmus
Improved Sobriety Test Procedures
as set forth by the National Highway Traffic Safety Administration

[Signature] *[Signature]*
Director, Division of Highway Traffic Safety Instructor

ORIGINAL COURSE DATES

	DATE	Refresher Course PLACE	INSTRUCTOR
1.	12/14/15	Sayreville PD	Cham
2.			
3.			

DEPARTMENT OF
Traffic and Public Safety
This is to certify that

Marcin Nierodzik
Dunellen Police

IS QUALIFIED AND COMPETENT TO CONDUCT OPERATIONS OF THE ALCOLTEST 7110 MKIII-C
THIS LAWS OF 1985 IN THE OPERATION OF THE ALCOLTEST 7110 MKIII-C
A METHOD TO EXTENDING INVESTIGATION
GIVEN UNDERMY HAND AT TRENTON, NEW JERSEY THIS 13th DAY OF April

[Signature] *[Signature]*
SUPERINTENDENT ATTORNEY GENERAL
NEW JERSEY STATE POLICE STATE OF NEW JERSEY