

Copy

# Calibrating Unit New Standard Solution Report

## Equipment

Alcotest 7110 MKIII-C	Serial No.:	ARTL-0009
Location: DUNELLEN POLICE DEPT.		
Calibration File No.: 01561	Calib. Date: 07/25/2022	Calib. No.: 00048
Certification File No.: 01562	Cert. Date: 07/25/2022	Cert. No.: 00039
Linearity File No.: 01563	Lin. Date: 07/25/2022	Lin. No.: 00040
Solution File No.: 01566	Soln. Date: 08/27/2022	Soln. No.: 00299
Sequential File No.: 01566	File Date: 08/27/2022	

Calibrating Unit: WET  
 Control Solution %: 0.100%  
 Solution Control Lot: 22070

Model No.: CU-34  
 Serial No.: DDUM S3-0213  
 Expires: 02/17/2024  
 Bottle No.: 1205

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	23:56D		
Control 1 EC	0.101%	23:57D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	23:57D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:57D		
Control 2 EC	0.101%	23:58D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.101%	23:58D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	23:59D		
Control 3 EC	0.101%	23:59D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.102%	23:59D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	00:00D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

### Changed By:

Last Name: GREEN

First Name: JOSEPH

MI: P

Signature: \_\_\_\_\_

Badge No.: 52

Date: 08/28/2022

<b>Certificate of Training</b>	
This is to certify that	
<b>Joseph P. Green</b>	
has successfully completed practitioner training in the use of	
<b>Horizontal Gaze Nystagmus</b>	
<b>Improved Sobriety Test Procedures</b>	
as set forth by the National Highway Traffic Safety Administration	
 E. Hunter <small>Director, Division of Highway Traffic Safety</small>	 <small>Instructor</small>