

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARTL-0009
Location:	DUNELLEN POLICE DEPT.	
Calibration File No.:	01387	Calib. Date: 06/15/2020
Certification File No.:	01388	Cert. Date: 06/15/2020
Linearity File No.:	01389	Lin. Date: 06/15/2020
Solution File No.:	01399	Soln. Date: 08/18/2020
Sequential File No.:	01399	File Date: 08/18/2020

Calibrating Unit: WET	Model No.: CU-34	Serial No.: DDUM S3-0218
Control Solution %: 0.100%		Expires: 07/06/2022
Solution Control Lot: 20320		Bottle No.: 1068

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	22:22D		
Control 1 EC	0.100%	22:23D	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	22:23D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:23D		
Control 2 EC	0.101%	22:24D	34.0°C	*** TEST PASSED ***
Control 2 IR	0.100%	22:24D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:25D		
Control 3 EC	0.101%	22:26D	34.0°C	*** TEST PASSED ***
Control 3 IR	0.101%	22:26D	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:26D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: NIERODZIK

First Name: MARCIN

MI: -

Signature: *[Signature]* #46

Badge No.: 46

Date: 08/18/2020

DEPARTMENT OF
Traffic and Public Safety
Helping to keep you safe

NEW JERSEY STATE POLICE

OFFICER: *[Signature]* #46

DATE: April 17th 2020

TIME: 17:00

LOCATION: DUNELLEN POLICE DEPT.

EQUIPMENT: Alcotest 7110 MKIII-C

SOLUTION CONTROL LOT: 20320

CONTROL SOLUTION %: 0.100%

CALIBRATION FILE NO.: 01387

CERTIFICATION FILE NO.: 01388

LINEARITY FILE NO.: 01389

SOLUTION FILE NO.: 01399

SEQUENTIAL FILE NO.: 01399

CALIBRATING UNIT: WET

CONTROL SOLUTION %: 0.100%

SOLUTION CONTROL LOT: 20320

CHIEF FORENSIC SCIENTIST: *[Signature]*

ORIGINAL COURSE DATES

DATE	Refresh Course PLACE	INSTRUCTOR
1. 12/17/15	Sayreville PD	<i>[Signature]</i>
2.		
3.		

Certificate of Training
This is to certify that

Marcin Nierodzik

has successfully completed practitioner training in the use of
Horizontal Gaze Nystagmus
Improved Sobriety Test Procedures
as set forth by the National Highway Traffic Safety Administration

[Signature]
Director, Division of Highway Traffic Safety

[Signature]
Instructor

printed on