

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARTL-0009
Location:	DUNELLEN POLICE DEPT.		
Calibration File No.:	01331	Calib. Date:	09/20/2019
Certification File No.:	01332	Cert. Date:	09/20/2019
Linearity File No.:	01333	Lin. Date:	09/20/2019
Solution File No.:	01365	Soln. Date:	01/25/2020
Sequential File No.:	01365	File Date:	01/25/2020

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDUM S3-0213
Control Solution %:	0.100%			Expires:	09/09/2021
Solution Control Lot:	19240			Bottle No.:	0938

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	05:20S		
Control 1 EC	0.099%	05:21S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.098%	05:21S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:22S		
Control 2 EC	0.099%	05:22S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.099%	05:22S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:23S		
Control 3 EC	0.099%	05:24S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	05:24S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	05:25S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NUSP) Chief Forensic Scientist.

Changed By:

Last Name: ROBBINS

First Name: BRIAN

MI: E

Signature: Sgt. B. Robbins #41

Badge No.: 41

Date: 01/25/2020