

# Calibrating Unit New Standard Solution Report

**Equipment** Alcotest 7110 MKIII-C  
**Location:** DUNELLEN POLICE DEPT.  
**Calibration File No.:** 01264  
**Certification File No.:** 01265  
**Linearity File No.:** 01266  
**Solution File No.:** 01296  
**Sequential File No.:** 01296

**Serial No.:** ARTL-0009  
**Calib. No.:** 00039  
**Cert. No.:** 00030  
**Lin. No.:** 00031  
**Soln. No.:** 00254

**Calib. Date:** 11/30/2018  
**Cert. Date:** 11/30/2018  
**Lin. Date:** 11/30/2018  
**Soln. Date:** 04/24/2019  
**File Date:** 04/24/2019

**Calibrating Unit:** WET  
**Control Solution %:** 0.100%  
**Solution Control Lot:** 19060

**Model No.:** CU-34  
**Serial No.:** DDUM S3-0218  
**Expires:** 02/11/2021  
**Bottle No.:** 0844

Function	Result	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	%BAC			
Control 1 EC	0.000%	22:42D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.101%	22:43D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:44D		
Control 2 EC	0.100%	22:44D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	22:44D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:45D		
Control 3 EC	0.101%	22:46D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	22:46D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	22:46D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**

**Last Name:** ROBBINS

**First Name:** BRIAN

**MI:** E

**Signature:**  #41

**Badge No.:** 41

**Date:** 04/24/2019