

Calibrating Unit

New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARTL-0009
Location:	DUNELLEN POLICE DEPT.		
Calibration File No.:	01303	Calib. Date:	05/20/2019
Certification File No.:	01304	Cert. No.:	00040
Linearity File No.:	01305	Cert. Date:	05/20/2019
Solution File No.:	01327	Lin. Date:	05/20/2019
Sequential File No.:	01327	Soln. Date:	08/21/2019
		Soln. No.:	00031
		File Date:	08/21/2019
		Lin. No.:	00032
		Soln. No.:	00258
Calibrating Unit:	WET	Model No.:	CU-34
Control Solution %:	0.100%	Serial No.:	DDUM S3-0218
Solution Control Lot:	19060	Expires:	02/11/2021
		Bottle No.:	0929

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	21:08D		
Control 1 EC	0.101%	21:08D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.102%	21:08D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:09D		
Control 2 EC	0.101%	21:10D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.102%	21:10D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:10D		
Control 3 EC	0.101%	21:11D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.102%	21:11D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:12D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: DELBUONO

First Name: JEFFREY

MI: M

Signature:



Badge No.: 32

Date: 08/21/2019

