

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C
Location: DUNELLEN POLICE DEPT.
Calibration File No.: 01264
Certification File No.: 01265
Linearity File No.: 01266
Solution File No.: 01279
Sequential File No.: 01279

Serial No.: ARTL-0009
Calib. No.: 00039
Cert. No.: 00030
Lin. No.: 00031
Soln. No.: 00252

Calib. Date: 11/30/2018
Cert. Date: 11/30/2018
Lin. Date: 11/30/2018
Soln. Date: 02/27/2019
File Date: 02/27/2019


Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 18290

Model No.: CU-34
Serial No.: DDUM S3-0218
Expires: 09/17/2020
Bottle No.: 1372

Function	Result	Time	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	%BAC 0.000%	20:08S		
Control 1 EC	0.099%	20:09S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.099%	20:09S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:09S		
Control 2 EC	0.100%	20:10S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.098%	20:10S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:11S		
Control 3 EC	0.100%	20:11S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	20:11S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	20:12S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By: 
Last Name: DOHERTY
Signature: _____

First Name: JOHN **MI:** J
Badge No.: 30
Date: 02/27/2019