

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.: ARTL-0009
Location:	DUNELLEN POLICE DEPT.	
Calibration File No.:	01303	Calib. Date: 05/20/2019
Certification File No.:	01304	Calib. No.: 00040
Linearity File No.:	01305	Cert. Date: 05/20/2019
Solution File No.:	01319	Cert. No.: 00031
Sequential File No.:	01319	Lin. Date: 05/20/2019
		Lin. No.: 00032
		Soln. Date: 07/22/2019
		Soln. No.: 00257
		File Date: 07/22/2019

Calibrating Unit: WET	Model No.: CU-34	Serial No.: DDUM S3-0218
Control Solution %: 0.100%		Expires: 02/11/2021
Solution Control Lot: 19060		Bottle No.: 0827

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	08:45D		
Control 1 EC	0.100%	08:46D	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	08:46D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:47D		
Control 2 EC	0.100%	08:47D	33.9°C	*** TEST PASSED ***
Control 2 IR	0.101%	08:47D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:48D		
Control 3 EC	0.101%	08:49D	33.9°C	*** TEST PASSED ***
Control 3 IR	0.101%	08:49D	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	08:49D		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: DELEUONO


First Name: JEFFREY

MI: M

Signature: 

Badge No.: 32




Date: 07/22/2019



JEFFREY DELEUONO
DUNELLEN POLICE

ALCOTEST 7110 MKIII-C

July

ORIGINAL COURSE DATES		
DATE	PLAC	INSTRUCTOR
2-24-11	SARASOTA PD	
5-7-13	JAYREVILLE	
1-13-15	SARASOTA PD	
7/6/17	JAYREVILLE PD	R.D. 7
6:		
6:		
7:		
6:		
6:		
6:		
6:		