

Calibrating Unit New Standard Solution Report

Equipment Alcotest 7110 MKIII-C
Location: DUNELLEN POLICE DEPT.
Calibration File No.: 01264
Certification File No.: 01265
Linearity File No.: 01266
Solution File No.: 01284
Sequential File No.: 01284

Serial No.: ARTL-0009

Calib. Date: 11/30/2018
Cert. Date: 11/30/2018
Lin. Date: 11/30/2018
Soln. Date: 03/27/2019
File Date: 03/27/2019

Calib. No.: 00039
Cert. No.: 00030
Lin. No.: 00031
Soln. No.: 00253

Calibrating Unit: WET
Control Solution %: 0.100%
Solution Control Lot: 18290

Model No.: CU-34
Serial No.: DDUM S3-0218
Expires: 09/17/2020
Bottle No.: 1366

Function	Result %BAC	Time HH:MM	Temperature Simulator (°C)	Comment(s) or Error(s)
Ambient Air Blank	0.000%	03:28S		
Control 1 EC	0.099%	03:29S	33.9°C	*** TEST PASSED ***
Control 1 IR	0.100%	03:29S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:30S		
Control 2 EC	0.100%	03:30S	33.9°C	*** TEST PASSED ***
Control 2 IR	0.100%	03:30S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:31S		
Control 3 EC	0.101%	03:32S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.100%	03:32S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	03:32S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: DELBUONO

First Name: JEFFREY

MI: M

Signature:



Badge No.: 32

Date: 03/27/2019