

Calibrating Unit New Standard Solution Report

Equipment	Alcotest 7110 MKIII-C	Serial No.:	ARTL-0009
Location:	DUNELLEN POLICE DEPT.		
Calibration File No.:	01331	Calib. Date:	09/20/2019
Certification File No.:	01332	Cert. Date:	09/20/2019
Linearity File No.:	01333	Lin. Date:	09/20/2019
Solution File No.:	01354	Soln. Date:	11/28/2019
Sequential File No.:	01354	File Date:	11/28/2019

Calibrating Unit:	WET	Model No.:	CU-34	Serial No.:	DDUM S3-0213
Control Solution %:	0.100%			Expires:	09/09/2021
Solution Control Lot:	19240			Bottle No.:	0999

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	21:39S		
Control 1 EC	0.098%	21:39S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.100%	21:39S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:40S		
Control 2 EC	0.098%	21:41S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.098%	21:41S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:41S		
Control 3 EC	0.098%	21:42S	34.0°C	*** TEST PASSED ***
Control 3 IR	0.099%	21:42S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	21:43S		

All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

Changed By:

Last Name: DELBUONO

First Name: JEFFREY

MI: M

Signature: _____



Badge No.: 32

Date: 11/28/2019




JEFFREY DELBUONO
DUNELLEN BOROUGH POLICE






IS QUALIFIED AND COMPETENT TO CONDUCT COURSE DATES AND IS PURSUANT TO CHAPTER 142 OF THE LAWS OF 1966 IN THE OPERATION OF THE ALCOLTEST 7110 MKIII-C A METHOD TO DETERMINE INTOXICATION.

GIVEN UNDER MY HAND AT TRENTON, NEW JERSEY, this _____ and _____ DAY OF July
TWO THOUSAND AND NINE


SUPERINTENDENT
NEW JERSEY STATE POLICE


ATTORNEY GENERAL
STATES OF NEW JERSEY

ORIGINAL COURSE DATES

DATE	Refresher Course PLACE	INSTRUCTOR
1. 2-24-11	SAYREVILLE PD	
2. 5-7-13	SAYREVILLE	
3. 1-13-15	SAYREVILLE PD	
4. 7/6/17	SAYREVILLE PD	
5. 11/26/17	Sayreville PD	
6.		
7.		
8.		
9.		

S.P. 2038 (Rev. 07/07)