

# Calibrating Unit New Standard Solution Report

<b>Equipment</b>	Alcotest 7110 MKIII-C	Serial No.: ARTL-0009
Location:	DUNELLEN POLICE DEPT.	
Calibration File No.:	01438	Calib. Date: 12/03/2020
Certification File No.:	01439	Cert. Date: 12/03/2020
Linearity File No.:	01440	Lin. Date: 12/03/2020
Solution File No.:	01459	Soln. Date: 02/28/2021
Sequential File No.:	01459	File Date: 02/28/2021


Calibrating Unit: WET	Model No.: CU-34	Serial No.: DDUM S3-0213
Control Solution %: 0.100%		Expires: 07/06/2022
Solution Control Lot: 20320		Bottle No.: 0936

Function	Result	Time	Temperature	Comment(s)
	%BAC	HH:MM	Simulator (°C)	or Error(s)
Ambient Air Blank	0.000%	04:25S		
Control 1 EC	0.100%	04:26S	34.0°C	*** TEST PASSED ***
Control 1 IR	0.101%	04:26S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:26S		
Control 2 EC	0.097%	04:27S	34.0°C	*** TEST PASSED ***
Control 2 IR	0.096%	04:27S	34.0°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:28S		
Control 3 EC	0.097%	04:29S	33.9°C	*** TEST PASSED ***
Control 3 IR	0.095%	04:29S	33.9°C	*** TEST PASSED ***
Ambient Air Blank	0.000%	04:29S		


All tests within acceptable tolerance.

On this date, I installed the above indicated "NEW SOLUTION" in accordance with Alcotest 7110 operator training and procedures established by the (NJSP) Chief Forensic Scientist.

**Changed By:**  
 Last Name: DELBUONO      First Name: JEFFREY      MI: M  
 Signature: *[Signature]*      Badge No.: 32      Date: 02/28/2021



DEPARTMENT OF  
**Traffic and Public Safety**  
 JEFFREY DELBUONO  
 DUNELLEN POLICE  
 IS QUALIFIED AND COMPETENT TO CONDUCT THE OPERATION OF THE ALCOLTEST 7110 MKIII-C  
 THE LAWS OF NJ IN THE OPERATION OF THE ALCOLTEST 7110 MKIII-C  
 A METHOD TO DETERMINE INTOXICATION  
 GIVEN UNDER MY NAME AT TRAINING ON 28 DAY OF July  
 TWO THOUSAND AND NINE



**Certificate of Training**  
 This is to certify that  
**Jeffrey Delbuono**  
 has successfully completed practitioner training for use of  
**Horizontal Gaze Nystagmus**  
 Improved Sobriety Test Protocol as  
 set forth by the National Highway Traffic Safety Administration

ORIGINAL COURSE DATES

DATE	TRAINING PLACE	INSTRUCTOR
2-24-11	SAYREVILLE PD	
5-7-13	SAYREVILLE	<i>[Signature]</i>
1-13-15	SAYREVILLE PD	
7-16-17	SAYREVILLE PD	<i>[Signature]</i>
11-26-17	Sayreville PD	<i>[Signature]</i>